

**LUDLUM MODEL 53 & 53B
PERSONNEL PORTAL
OPERATOR'S MANUAL**

January 2023 Revision 4

Version 1.4.2 / 21505N01

LUDLUM MODEL 53 & 53B PERSONNEL PORTAL OPERATOR'S MANUAL

January 2023 Revision 4

Version 1.4.2 / 21505N01



LUDLUM MEASUREMENTS, INC.
501 OAK STREET, P.O. BOX 810
SWEETWATER, TEXAS 79556
325-235-5394, FAX: 325-235-4672

STATEMENT OF WARRANTY

Ludlum Measurements, Inc. warrants the products covered in this manual to be free of defects due to workmanship, material, and design for a period of twelve months from the date of delivery. The calibration of a product is warranted to be within its specified accuracy limits at the time of shipment. In the event of instrument failure, notify Ludlum Measurements to determine if repair, recalibration, or replacement is required.

This warranty excludes the replacement of photomultiplier tubes, G-M and proportional tubes, and scintillation crystals which are broken due to excessive physical abuse or used for purposes other than intended.

There are no warranties, express or implied, including without limitation any implied warranty of merchantability or fitness, which extend beyond the description of the face there of. If the product does not perform as warranted herein, purchaser's sole remedy shall be repair or replacement, at the option of Ludlum Measurements. In no event will Ludlum Measurements be liable for damages, lost revenue, lost wages, or any other incidental or consequential damages, arising from the purchase, use, or inability to use product.

RETURN OF GOODS TO MANUFACTURER

If equipment needs to be returned to Ludlum Measurements, Inc. for repair or calibration, please send to the address below. All shipments should include documentation containing return shipping address, customer name, telephone number, description of service requested, and all other necessary information. Your cooperation will expedite the return of your equipment.

**LUDLUM MEASUREMENTS, INC.
ATTN: REPAIR DEPARTMENT
501 OAK STREET
SWEETWATER, TX 79556**

**800-622-0828 325-235-5494
FAX 325-235-4672**

Table of Contents

INTRODUCTION	1
FEATURES.....	3
SOFTWARE LICENSE AGREEMENT.....	4
SAFETY CONSIDERATIONS	6
ENVIRONMENTAL CONDITIONS FOR NORMAL USE.....	6
REPLACEMENT OF FUSES.....	6
CLEANING PRECAUTIONS	6
LEAD SHIELDED DETECTORS	6
OPERATION.....	7
CONTROLS AND FUNCTIONS.....	7
POWER ON SEQUENCE.....	7
POWERING THE MODEL 53 ON	7
POWERING THE MODEL 53 OFF.....	8
STATUS COLORS	8
INSTRUMENT FAILURE	9
SUM ZONES	9
SUM ZONE ALARM PRIORITY	9
NORMAL USER OPERATION BY USER MODE.....	10
SUPERVISOR	16
OPERATION	16
OPERATE SCREEN	16
DETECTOR DETAIL SCREEN	18
SHUTDOWN MENU	19
SETUP SCREEN	20
RADIOLOGICAL SETUP	23
OCCUPANCY MODES	23
<i>Walk-Thru Mode</i>	23
<i>Pause Mode</i>	24
<i>Pause-And-Turn Mode</i>	24
<i>Front-And-Back Mode</i>	24
COUNTING MODES	24
<i>Mode 1 – Maximum Sensitivity</i>	24
<i>Mode 2 – Fixed MDA</i>	26
<i>Mode 3 – Minimum Count Time</i>	26
<i>Count Mode Settings</i>	28
FAST CLEAN.....	29
FAST ALARM.....	30
INDIVIDUAL ALARMS	30
UNITS OF MEASURE	31
UPDATING BACKGROUND	31
<i>Background Update Method</i>	32
BACKGROUND ALARMS	33
CALCULATIONS	33

OPERATIONAL SETUP.....	35
GENERAL	35
<i>Serial Number</i>	35
<i>Site</i>	36
<i>Location</i>	36
<i>Customer ID</i>	36
<i>Password</i>	36
<i>Alarm Hold</i>	36
<i>Infrared Delay</i>	36
<i>Infrared Timeout</i>	36
<i>Enable Check-out Mode</i>	36
OPTIONS.....	37
<i>Require Employee ID to Start Count</i>	37
<i>Require Password to Clear Alarms</i>	37
<i>Enable Automatic Recount</i>	37
<i>Enable Incomplete Alarm</i>	37
<i>Show Alarm Result</i>	38
<i>Show Counts as Activity</i>	38
<i>Voltage Cal Enabled</i>	38
<i>Latch Failures</i>	38
<i>Ignore Approach/Exit Sensors</i>	38
LANGUAGE.....	39
LOGGING.....	40
COM PORT.....	41
NETWORK	42
DETECTORS	43
BACKGROUND.....	43
HIGH VOLTAGE (HV).....	45
LOWER LEVEL DISCRIMINATOR (LLD).....	46
VOLTAGE CALIBRATION	46
<i>HV Actual Calibration Constant</i>	47
<i>HV Read Back Calibration Constant</i>	49
<i>LLD Actual Calibration Constant</i>	49
DETECTOR SERIAL NUMBERS.....	50
SOURCE CHECK.....	52
FALSE ALARM	54
BUTTONS/RELAYS.....	56
VARIANCE/MEAN	57
SCALER.....	59
FOM-HV	61
SELECT DETECTORS FOR HV PLATEAU	61
COUNT TIME SETTINGS	62
SOURCE SIZE.....	62
HV SETTINGS.....	62
RUNNING THE FOM-HV	62
EFFICIENCIES.....	64

CONFIGURATION	64
SOURCES	65
ISOTOPIC MIX.....	66
EFFICIENCIES	67
COMPOSITE EFFICIENCY.....	67
VIEW LOGS	68
CHANGES	70
RECYCLING.....	73

Section

1

Introduction

The Model 53 Gamma Personnel Portal is used to detect gamma radiation in or on personnel passing through the portal. The portal has eight large plastic scintillation detectors and is shielded with lead, making it very heavy and highly sensitive. The Model 53 is designed to be user friendly – users follow automated voice prompts and see status clearly on the 20.3 cm (8 in.) color liquid crystal display (LCD). Instrument technicians have password-protected access to advanced automated routines to calibrate or verify operation.

Another variation of this personnel portal is the Model 53B, which detects beta radiation in addition to gamma. The detector access doors on the sides and on top have openings with metallized polyester windows to permit beta transmission. The window thickness is 1.2 mg/cm². Beta efficiency for ⁹⁰Sr/Y is ≥ 7% at 7.6 cm (3 in.) from the side and top detectors. It is 0% on the foot detectors where there are no openings.

The Model 53 Gamma Personnel Portal can be configured with several different options including:

- ☼ 1 or 2 inch thick lead shielding
- ☼ 1 or 2 light stack status indicators

There are three statistical counting modes to maximize throughput, sensitivity, or to fix the MDA (Minimum Detectable Activity). Several parameters can be modified to adjust the alarm set point, including the false alarm probability, detection probability, background sigma coefficient (K_B), and the composite sigma coefficient (K_{S+B}). Fast alarm and clean options provide the ability to quickly determine if personnel are contaminated or clean before the entire count cycle has ended.

There are also several user modes to choose from. Calibrators can choose to use the walk-through mode, a pause mode, a pause-and-turn mode, or front-and-back mode. Voice prompts may be customized in any language for each of these modes. These prompts can, for example, dictate in-house procedures to follow.

Instrument status is communicated clearly through the use of color LCD displays, light stack status indicators, and the Ethernet link. Log, setup, and report files can easily be transferred to/from the Model 53 via a local network or a USB flash drive.

Section

2

Features

- ☢ Large plastic scintillator detectors for optimized detection efficiency
- ☢ Fast Alarm/Fast Clean counting technology for shorter counting cycles
- ☢ Three counting modes to maximize throughput, maximize sensitivity, or fix the MDA
- ☢ Automatic background updating
- ☢ Recount and contaminated detector check options
- ☢ Individual detector, sum zone, and sum channel alarms
- ☢ PIR (Passive Infrared) sensor to sense personnel approach
- ☢ Photoelectric sensor to detect personnel presence
- ☢ Two levels of password security
- ☢ Four user modes: Walk-Thru, Pause, Pause-and Turn, Front-and-Back
- ☢ 1-inch thick lead shielding (option for 2 inch)
- ☢ 196 x 61 x 50.8 cm (77 x 24 x 20 in.) inside dimensions
- ☢ Single Board Computer (SBC) running Windows 10 Pro Operating System
- ☢ Additional USB (Universal Serial Bus) connectors to transfer information from the system, or for input of user ID, etc.

Section

3

Software License Agreement

BY INSTALLING THIS SOFTWARE, YOU ARE CONSENTING TO BE BOUND BY THIS AGREEMENT. IF YOU DO NOT AGREE TO ALL OF THE TERMS OF THIS AGREEMENT, DO NOT INSTALL THE PRODUCT.

Single User License Grant: Ludlum Measurements, Inc. ("Ludlum") and its suppliers grant to Customer ("Customer") a nonexclusive and nontransferable license to use the Ludlum software ("Software") in object code form solely on a single central processing unit owned or leased by Customer or otherwise embedded in equipment provided by Ludlum.

Customer may make one (1) archival copy of the Software provided Customer affixes to such copy all copyright, confidentiality, and proprietary notices that appear on the original.

EXCEPT AS EXPRESSLY AUTHORIZED ABOVE, CUSTOMER SHALL NOT: COPY, IN WHOLE OR IN PART, SOFTWARE OR DOCUMENTATION; MODIFY THE SOFTWARE; REVERSE COMPILE OR REVERSE ASSEMBLE ALL OR ANY PORTION OF THE SOFTWARE; OR RENT, LEASE, DISTRIBUTE, SELL, OR CREATE DERIVATIVE WORKS OF THE SOFTWARE.

Customer agrees that aspects of the licensed materials, including the specific design and structure of individual programs, constitute trade secrets and/or copyrighted material of Ludlum. Customer agrees not to disclose, provide, or otherwise make available such trade secrets or copyrighted material in any form to any third party without the prior written consent of Ludlum. Customer agrees to implement reasonable security measures to protect such trade secrets and copyrighted material. Title to Software and documentation shall remain solely with Ludlum.

LIMITED WARRANTY. Ludlum warrants that for a period of ninety (90) days from the date of shipment from Ludlum: (i) the media on which the Software is furnished will be free of defects in materials and workmanship under normal use; and (ii) the Software substantially conforms to its published specifications. Except for the foregoing, the Software is provided AS IS. This limited warranty extends only to Customer as the original licensee. Customer's exclusive remedy and the entire liability of Ludlum and its suppliers under this limited warranty will be, at Ludlum or its service center's option, repair, replacement, or refund of the Software if reported (or, upon request, returned) to the party supplying the Software to Customer. In no event does Ludlum warrant that the Software is error free or that Customer will be able to operate the Software without problems or interruptions.

This warranty does not apply if the software (a) has been altered, except by Ludlum, (b) has not been installed, operated, repaired, or maintained in accordance with instructions supplied by Ludlum, (c) has been subjected to abnormal physical or electrical stress, misuse, negligence, or accident, or (d) is used in ultra hazardous activities.

DISCLAIMER. EXCEPT AS SPECIFIED IN THIS WARRANTY, ALL EXPRESS OR IMPLIED

CONDITIONS, REPRESENTATIONS, AND WARRANTIES INCLUDING, WITHOUT LIMITATION, ANY IMPLIED WARRANTY OF MERCHANTABILITY, FITNESS FOR A PARTICULAR PURPOSE, NONINFRINGEMENT OR ARISING FROM A COURSE OF DEALING, USAGE, OR TRADE PRACTICE, ARE HEREBY EXCLUDED TO THE EXTENT ALLOWED BY APPLICABLE LAW. IN NO EVENT WILL LUDLUM OR ITS SUPPLIERS BE LIABLE FOR ANY LOST REVENUE, PROFIT, OR DATA, OR FOR SPECIAL, INDIRECT, CONSEQUENTIAL, INCIDENTAL, OR PUNITIVE DAMAGES HOWEVER CAUSED AND REGARDLESS OF THE THEORY OF LIABILITY ARISING OUT OF THE USE OF OR INABILITY TO USE THE SOFTWARE EVEN IF LUDLUM OR ITS SUPPLIERS HAVE BEEN ADVISED OF THE POSSIBILITY OF SUCH.

DAMAGES. In no event shall Ludlum's or its suppliers' liability to Customer, whether in contract, tort (including negligence), or otherwise, exceed the price paid by Customer. The foregoing limitations shall apply even if the above-stated warranty fails of its essential purpose. **SOME STATES DO NOT ALLOW LIMITATION OR EXCLUSION OF LIABILITY FOR CONSEQUENTIAL OR INCIDENTAL DAMAGES.**

The above warranty DOES NOT apply to any beta software, any software made available for testing or demonstration purposes, any temporary software modules or any software for which Ludlum does not receive a license fee. All such software products are provided AS IS without any warranty whatsoever.

This License is effective until terminated. Customer may terminate this License at any time by destroying all copies of Software including any documentation. This License will terminate immediately without notice from Ludlum if Customer fails to comply with any provision of this License. Upon termination, Customer must destroy all copies of Software.

Software, including technical data, is subject to U.S. export control laws, including the U.S. Export Administration Act and its associated regulations, and may be subject to export or import regulations in other countries. Customer agrees to comply strictly with all such regulations and acknowledges that it has the responsibility to obtain licenses to export, re-export, or import Software.

This License shall be governed by and construed in accordance with the laws of the State of Texas, United States of America, as if performed wholly within the state and without giving effect to the principles of conflict of law. If any portion hereof is found to be void or unenforceable, the remaining provisions of this License shall remain in full force and effect. This License constitutes the entire License between the parties with respect to the use of the Software.

Restricted Rights - Ludlum's software is provided to non-DOD agencies with RESTRICTED RIGHTS and its supporting documentation is provided with LIMITED RIGHTS. Use, duplication, or disclosure by the Government is subject to the restrictions as set forth in subparagraph "C" of the Commercial Computer Software - Restricted Rights clause at FAR 52.227-19. In the event the sale is to a DOD agency, the government's rights in software, supporting documentation, and technical data are governed by the restrictions in the Technical Data Commercial Items clause at DFARS 252.227-7015 and DFARS 227.7202. Manufacturer is Ludlum Measurements, Inc. 501 Oak Street Sweetwater, Texas 79556.

Section

4

Safety Considerations

Environmental Conditions for Normal Use

Indoor use only

No maximum altitude

Temperature range of to 0 to 50 °C (32 to 122 °F)

Maximum relative humidity of less than 95% (non-condensing)

Mains supply voltage range 85-250 Vac, 50-60 Hz single phase (less than 1 A) to desktop power supply supplies DC voltage to instrument.

Replacement of Fuses

Warning!

For continued protection against risk of fire, replace only with fuses of the specified type and current rating!

Cleaning Precautions

The Model 53 is ruggedly designed and requires very little maintenance. The only routine maintenance item is the periodic cleaning of the displays and sensors using a soft cloth dampened with water only.

Lead Shielded Detectors

The detectors are shielded with lead, which makes the system very heavy. Use caution when moving the portal. Lead shielding can and should be removed to facilitate moving.

Caution!

Stabilizer feet should not be used to level device. They are only to stabilize device!

Section

5

Operation

Controls and Functions

The following is a description of the controls and functions on the Gamma Personnel Portal.

LCDs (Liquid Crystal Displays): 20.3 cm (8 in.) color LCDs are mounted on ingress and egress sides

Acknowledge Button(s): Red alarm acknowledge buttons are mounted on the ingress and egress sides. Press button once to acknowledge the audio and again to clear the alarm.

Detector Doors: Locking panel doors provide internal access to the power switch, main electronics, and detectors.

Light Stack with Audio: Three color (red, yellow, and green) lights indicate the Gamma Personnel Portal status.

- Red- Radiological Alarm
- Yellow- Inoperational Warning
- Green- Unit Operational

Power on Sequence

When the Gamma Personnel Portal is first powered up, approximately one to two minutes are required to finish booting, and display the initializing screen. After initialization, the system will determine the level of background radiation. The system then begins updating background radiation levels every second.

Powering the Model 53 On

1. Using the supplied key, unlock and open the detector panel.
2. Connect the AC adapter to a 120-240 Vac electrical outlet.

3. Close and lock the panel.
4. Turn the Start/Stop key switch fully clockwise momentarily and release.
5. The Model 53 will take approximately one to two minutes to start.

Powering the Model 53 Off

1. Tap the Exit button on the main screen and enter the level 1 or 2 password.
2. Tap the Shutdown Gamma Personnel Portal button.
3. Wait until the system powers down completely.
4. Alternately, the Start/Stop key switch can be turned fully clockwise momentarily and released to initiate Model 53 system shutdown.

Note: Do not remove power from the Model 53 without first shutting down the computer.

Status Colors

The Gamma Personnel Portal uses color to identify the status of the instrument at any given time. The status box in the upper-left corner of the main screen and the detector map change colors depending on the various states of the Gamma Personnel Portal and detectors. The colors are:

-  Green – Indicates the Gamma Personnel Portal is ready to scan personnel.
-  Yellow – Indicates the Gamma Personnel Portal has a failure.
-  Red – Indicates the Gamma Personnel Portal has a radiation alarm.
-  Orange – Indicates an incomplete count has occurred on the Gamma Personnel Portal.
-  White – Indicates the Gamma Personnel Portal is scanning personnel.
-  Blue – Indicates the Gamma Personnel Portal is initializing or updating background.

The lights and audio on the light stack operate as follows:

-  Red – Indicates the Gamma Personnel Portal has a radiation alarm. In addition to the voice audible alert, there will be a fast beeping audible alarm.

- ☢ Yellow – Indicates the Gamma Personnel Portal is in a fail condition or is not ready to scan personnel. If a fail condition is present, there will be a slow-beeping audible alarm. While in any setup screens, when the Gamma Personnel Portal is forcing a background update, or when the Supervisor application is not running, the yellow light will be ON.
- ☢ Green – Indicates the Gamma Personnel Portal is ready to scan personnel.

Instrument Failure

There are several conditions that cause system failure:

1. **LOW BACKGROUND** - The instrument continually checks the detectors for abnormally low readings as defined by the low-background alarm set points. When this condition is detected, the detector map on the main screen will indicate which detector failed. This condition normally signals a failure of either the detector or its associated preamp electronics.
2. **HIGH BACKGROUND** - The instrument continually checks the detectors for abnormally high readings as defined by the high-background alarm set points. When this condition is detected, the detector map on the main screen will indicate which detector failed. This condition normally signals a failure of either the detector or its associated preamp electronics, but can also warn of some nearby strong source of radiation. In either case, the possibility of such a nearby strong source of radiation warrants further investigation.

If the background for a detector fails or it becomes contaminated, the Scaler screen in the Setup menu can be used to determine if the failure or contamination has been corrected.

Sum Zones

The Model 53 can be set to use individual detectors or a combination of sum zones. The Gamma Personnel Portal physical detectors are grouped into various sum zones composed of two and three detector combinations. The sum zones are helpful in detection of widespread contamination.

Sum Zone Alarm Priority

Alarm priority is dependent on the number of detectors. A two-detector sum alarm will not be posted if either of its constituent detectors have alarmed. A three-detector sum alarm will not be posted if any of its single detectors or two-detector sum zones have alarmed.

Normal User Operation by User Mode

The Model 53 may be configured to use one of several user modes. The different user modes have different sensitivity levels and throughput levels. The user modes and operation are described below:

Walk-Thru Mode-user simply walks through the portal, achieving the highest throughput, but the lowest sensitivity. The three count modes (Max Sensitivity, Fixed MDA, and Min Count Time do not apply). Specific actions are:

- 1) User approaches portal if green status light is active. Portal is continually determining background levels of radiation.
- 2) User will trigger approach sensors, halting background measurements, and trigger an ENTRANCE vocal message, which is typically "PLEASE PASS THROUGH THE PORTAL." The green status light will flash slowly when only approach sensors are active.
- 3) When user interrupts the occupancy sensor in the center of the portal, the portal will look for radiation levels above the background measurements. The green status light will flash quickly while the portal is occupied.
- 4) If a radiological alarm is determined, the ALARM vocal message will be triggered, typically "RADIATION ALARM." If the automatic recount feature is enabled, the vocal messages "RETURN FOR RECOUNT" and "RECOUNTING" are used, depending on whether the user has cleared the occupancy sensor. If latching alarm mode is enabled, the RADIATION ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message "PLEASE CLEAR THE PORTAL" is then utilized until occupancy and approach sensors are clear.
- 5) If no alarm is triggered, the OK vocal message is activated once. The green status light flashes slowly until all sensors are clear. A preset delay, typically four seconds, follows. If any of the sensors are active after this time, the vocal message "PLEASE CLEAR THE PORTAL" is then utilized until occupancy and approach sensors are clear.
- 6) The green status light is activated (no flashing), and the portal will update the background measurements and be ready for additional occupancies.

Pause Mode-user pauses inside the portal until counting is complete, achieving lower throughput, but higher sensitivity than the Walk-Thru Mode. Specific actions are:

- 1) User approaches portal if green status light is active. Portal is continually determining background levels of radiation.
- 2) User will trigger approach sensors, halting background measurements, and trigger an ENTRANCE vocal message, which is typically “PLEASE ENTER AND PAUSE INSIDE THE PORTAL.” The green status light will flash slowly when only approach sensors are active.
- 3) When user interrupts the occupancy sensor in the center of the portal, the portal will look for radiation levels above the background measurements. The vocal PAUSE message, typically “PAUSE UNTIL COUNTING IS FINISHED,” is triggered. The green status light will flash quickly while the portal is occupied.
- 4) If the occupancy sensor is cleared before counting is complete, the vocal INCOMPLETE message is triggered, typically “COUNTING INCOMPLETE, PLEASE RETURN TO CENTER OF PORTAL.” The message is repeated for a preset time or until the occupancy sensor is triggered again. If the occupancy sensor is not triggered again and the preset time elapses, the portal will either:
 - Return to ready mode (steady green status light and background measurements) if all sensors are clear;
 - Activate the CLEAR PORTAL vocal message, typically “PLEASE CLEAR THE PORTAL,” if the approach sensors are not clear; or
 - Activate the INCOMPLETE ALARM message, typically “INCOMPLETE COUNT ALARM,” if the incomplete alarm option is selected during calibration. If latching alarm mode is enabled, the INCOMPLETE COUNT ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message “PLEASE CLEAR THE PORTAL” is then utilized until occupancy and approach sensors are clear.

- 5) If a radiological alarm is determined, the ALARM vocal message will be triggered, typically “RADIATION ALARM.” If the automatic recount feature is enabled, the vocal messages “RETURN FOR RECOUNT” and “RECOUNTING” are used, depending on whether the user has cleared the occupancy sensor. If latching alarm mode is enabled, the RADIATION ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message “PLEASE CLEAR THE PORTAL” is then utilized until occupancy and approach sensors are clear.
- 6) If no alarm is triggered and counting is complete, the OK vocal message, typically “OK PLEASE PROCEED,” is activated once. The green status light flashes slowly until all sensors are clear. A preset delay, typically four seconds, follows. If any of the sensors are active after this time, the vocal message “PLEASE CLEAR THE PORTAL” is then utilized until occupancy and approach sensors are clear.
- 7) The green status light is activated (no flashing), and the portal will update the background measurements and be ready for additional occupancies.

Pause-And-Turn Mode-user pauses inside the portal until counting is complete, then turns 90 degrees and counts again, achieving lower throughput, but higher sensitivity than the Pause Mode. Specific actions are:

- 1) User approaches portal if green status light is active. Portal is continually determining background levels of radiation.
- 2) User will trigger approach sensors, halting background measurements, and trigger an ENTRANCE vocal message, which is typically “PLEASE ENTER AND PAUSE INSIDE THE PORTAL.” The green status light will flash slowly when only approach sensors are active.
- 3) When the user interrupts the occupancy sensor in the center of the portal, the portal will look for radiation levels above the background measurements. The vocal PAUSE message, typically “PAUSE UNTIL COUNTING IS FINISHED,” is triggered. The green status light will flash quickly while the portal is occupied.
- 4) If the occupancy sensor is cleared before counting is complete, the vocal INCOMPLETE message is triggered, typically “COUNTING

INCOMPLETE, PLEASE RETURN TO CENTER OF PORTAL.” The message is repeated for a preset time or until the occupancy sensor is triggered again. If the occupancy sensor is not triggered again and the preset time elapses, the portal will either:

- return to ready mode (steady green status light and background measurements) if all sensors are clear;
 - activate the CLEAR PORTAL vocal message, typically “PLEASE CLEAR THE PORTAL,” if the approach sensors are not clear; or
 - activate the INCOMPLETE ALARM message, typically “INCOMPLETE COUNT ALARM,” if the incomplete alarm option is selected during calibration. If latching alarm mode is enabled, the INCOMPLETE COUNT ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message “PLEASE CLEAR THE PORTAL” is then utilized until occupancy and approach sensors are clear.
- 5) If a radiological alarm is determined, the ALARM vocal message will be triggered, typically “RADIATION ALARM.” If the automatic recount feature is enabled, the vocal messages “RETURN FOR RECOUNT” and “RECOUNTING” are used depending on whether the user has cleared the occupancy sensor. If latching alarm mode is enabled, the RADIATION ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message “PLEASE CLEAR THE PORTAL” is then utilized until occupancy and approach sensors are clear.
- 6) If no alarm is triggered and counting is complete, the OK-TURN vocal message, typically “OK PLEASE TURN LEFT AND REMAIN IN PORTAL...COUNTING,” is activated once. Return to Step 4, except when counting is completed without triggering an alarm. The OK vocal message, typically “OK PLEASE PROCEED,” is activated once. The green status light flashes slowly until all sensors are clear. A preset delay, typically four seconds, follows. If any of the sensors are active after this time, the vocal message “PLEASE CLEAR THE PORTAL” is then utilized until occupancy and approach sensors are clear.

- 7) The green status light is activated (no flashing), and the portal will update the background measurements and be ready for additional occupancies.

Front-And-Back Mode-user enters the portal, turns 90 degrees, steps forward until that user is against the detectors facing him or her and pauses until counting is complete, then steps backward until against the detectors behind him or her and counts again, achieving lower throughput, but higher sensitivity than the Pause Mode. Specific actions are:

- 8) User approaches portal if green status light is active. Portal is continually determining background levels of radiation.
- 9) User will trigger approach sensors, halting background measurements, and trigger an ENTRANCE vocal message, which is typically “PLEASE ENTER AND PAUSE INSIDE THE PORTAL.” The green status light will flash slowly when only approach sensors are active.
- 10) When the user interrupts the occupancy sensor in the center of the portal, an additional positioning vocal message is triggered, typically “PLEASE TURN LEFT, STEP FORWARD AND STAND AGAINST THE DETECTORS” is triggered. The portal then gives the “COUNTING” vocal message and begins looking for radiation levels above normal background. The green status light will flash quickly while the portal is occupied.
- 11) If the occupancy sensor is cleared before counting is complete, the vocal INCOMPLETE message is triggered, typically “COUNTING INCOMPLETE, PLEASE RETURN TO CENTER OF PORTAL.” The message is repeated for a preset time or until the occupancy sensor is triggered again. If the occupancy sensor is not triggered again and the preset time elapses, the portal will either:
 - return to ready mode (steady green status light and background measurements) if all sensors are clear;
 - activate the CLEAR PORTAL vocal message, typically “PLEASE CLEAR THE PORTAL,” if the approach sensors are not clear; or
 - activate the INCOMPLETE ALARM message, typically “INCOMPLETE COUNT ALARM,” if the incomplete alarm option is selected during calibration. If latching alarm

mode is enabled, the INCOMPLETE COUNT ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message "PLEASE CLEAR THE PORTAL" is then utilized until occupancy and approach sensors are clear.

- 12) If a radiological alarm is determined, the ALARM vocal message will be triggered, typically "RADIATION ALARM." If the automatic recount feature is enabled, the vocal messages "RETURN FOR RECOUNT" and "RECOUNTING" are used depending on whether the user has cleared the occupancy sensor. If latching alarm mode is enabled, the RADIATION ALARM status is maintained until the ALARM ACK button is pressed. If non-latching alarm mode is enabled, the alarm will sound for a preset time interval before clearing. In either case, the vocal message "PLEASE CLEAR THE PORTAL" is then utilized until occupancy and approach sensors are clear.
- 13) If no alarm is triggered and counting is complete, the "PLEASE STEP BACKWARD, AND STAND AGAINST THE DETECTORS," vocal message followed by "COUNTING" vocal message is issued. Return to Step 4, except when counting is completed without triggering an alarm. The OK vocal message, typically "OK PLEASE PROCEED," is activated once. The green status light flashes slowly until all sensors are clear. A preset delay, typically four seconds, follows. If any of the sensors are active after this time, the vocal message "PLEASE CLEAR THE PORTAL" is then utilized until occupancy and approach sensors are clear.
- 14) The green status light is activated (no flashing), and the portal will update the background measurements and be ready for additional occupancies.

Section

6

Supervisor

The Model 53 Gamma Personnel Portal is controlled by the Supervisor application. The Supervisor is started automatically when the Gamma Personnel Portal is booted. If the Supervisor is not running, it can be started from the Model 53 Gamma Personnel Portal icon on the desktop. When the Supervisor is not running, the yellow light on the light stack will be lit to indicate that the Gamma Personnel Portal is not in service.

Operation

For normal operation, the Supervisor requires little to no interaction. When interaction is required, the Supervisor can be controlled using an external USB keyboard and mouse. Numeric fields have up and down arrows, which provide a means to quickly increment or decrement the value. By clicking directly in a numeric field, an on-screen keypad is displayed for quick data entry.

Operate Screen

The Operate screen (see page 17 for screen shot) is the main display of the Gamma Personnel Portal and contains the following items:

Title Bar – Displays the user-definable site, location, customer ID, and LMI serial number of the Gamma Personnel Portal along the top of the screen.

Status Display – Displays the current status of the Gamma Personnel Portal. The status can be one of the following:

-  Ready
-  Clean
-  Counting
-  Alarm
-  High Alarm
-  Failure

- ☢ Updating Background
- ☢ Extending Background
- ☢ Initializing
- ☢ Contamination Check
- ☢ Contaminated

Status Text – Displays information about the current status, including which detector has alarmed or failed.


Clock/Timer – Displays the current time when not counting and a countdown timer when the Gamma Personnel Portal is counting or initializing a new background.

Detector Status Display – Graphical representation of the detectors with color-coded status display. Tapping on a detector brings up the detector detail screen.

Exit Button – Allows the user to exit the Supervisor application, restart the Gamma Personnel Portal, or shut down the Gamma Personnel Portal. It is password protected.

Setup Button – Allows access to the various setup screens. It is password protected.

Version – The software version and host board firmware version are displayed next to the LMI logo in the bottom, left corner of the screen.



Detector Detail Screen

A detail screen for the detectors can be displayed by clicking on any detector in the graphical display. The detail screen provides the background count, net count (if a count is in progress), alarm set point, and status. Also displayed are the values for each sum zone that the detector is included in.


Detector Detail						
Detector	Background (cps)	Net Count (cps)	Net Alarm (cps)	Net Fast Alarm (cps)	Net Fast Clean (cps)	Status
1	1,000	0	91	N/A	N/A	No Alarm
2	1,000	0	91	N/A	N/A	No Alarm
3	1,000	0	91	N/A	N/A	No Alarm
4	1,000	0	91	N/A	N/A	No Alarm
5	1,000	0	91	N/A	N/A	No Alarm
6	1,000	0	91	N/A	N/A	No Alarm
7	1,000	0	91	N/A	N/A	No Alarm
8	1,000	0	91	N/A	N/A	No Alarm
1+2	2,000	0	129	N/A	N/A	No Alarm
2+3	2,000	0	129	N/A	N/A	No Alarm
3+4	2,000	0	129	N/A	N/A	No Alarm
4+5	2,000	0	129	N/A	N/A	No Alarm
5+6	2,000	0	129	N/A	N/A	No Alarm
6+7	2,000	0	129	N/A	N/A	No Alarm
7+8	2,000	0	129	N/A	N/A	No Alarm
8+1	2,000	0	129	N/A	N/A	No Alarm
8+2+1	3,000	0	157	N/A	N/A	No Alarm
2+3+4	3,000	0	157	N/A	N/A	No Alarm
4+5+6	3,000	0	157	N/A	N/A	No Alarm
6+7+8	3,000	0	157	N/A	N/A	No Alarm

When a single detector alarms, no Sum Zone that includes that detector is allowed to alarm.

Close

Shutdown Menu

The Shutdown menu provides access to restart or shut down the Gamma Personnel Portal or to exit the Supervisor application and is accessed through the Exit button on the Operate screen. The level 1 or 2 password is required before accessing this screen. Always shut down the Gamma Personnel Portal properly through software or the key switch before removing power from the personnel portal monitor; otherwise, data corruption may occur.



Exit Supervisor – Closes the Supervisor application and provides access to the Windows 10 Embedded environment. This is necessary for performing software

updates, configuring windows settings, such as printing or networking, or to provide access to the data files. A confirmation window will be displayed before the Supervisor application is closed.

Restart Gamma Personnel Portal – This function performs a restart of the Gamma Personnel Portal's single board computer. A confirmation window will be displayed before the Gamma Personnel Portal is restarted.

Shutdown Gamma Personnel Portal – This function shuts down the Gamma Personnel Portal. This or the key switch option should always be used before removing power from the Gamma Personnel Portal to prevent data corruption.

Setup Screen


The Setup button on the Operate screen (see screen shot on page 21) provides access to the various setup and calibration functions of the Gamma Personnel Portal. The level 2 password provides full access to all setup screens. The level 1 password only provides access to the following screens:

- ☢ Save Setup Report
- ☢ Save Cal Report
- ☢ Update Background
- ☢ Source Check
- ☢ Buttons/Relays
- ☢ Scaler
- ☢ Reset Host
- ☢ View Log

If the Gamma Personnel Portal is in a fail or contaminated state (other than a host communication hardware failure), only the following setup screens are available regardless of the password level:

- ☢ Radiological Setup
- ☢ Operational Setup
- ☢ Detectors
- ☢ Save Setup Report

- ☢ Scaler
- ☢ Reset Host
- ☢ View Log



Radiological – Sets up the counting mode and its parameters, isotopic mix, occupancy mode, and units of measurement.

Operational – Sets up passwords, user-definable information, and options that control how the Supervisor application operates and displays information.

Detectors – Sets up the detector high voltage (HV) and lower level discriminator (LLD). A calibration wizard is available to calibrate the various voltages and calibration constants.

Save Setup Report – Saves a report that contains the current hardware and software configuration.

Save Cal Report – Saves a calibration report that contains all the calibration and setup information. The report covers the counting parameters, source information, detector settings, FOM data, and application settings. After saving the file, the report can be copied from the Gamma Personnel Portal and printed if necessary. Open the file in WordPad and change all margins to 0.5 inches and the font size from 10 pts to 9.

Update – Starts a new background initialization.

Source Check – Performs a source check to verify all detectors are operating correctly.

False Alarm – Performs a false alarm test with a specified number of counting cycles for each detector.

Buttons/Relays – Tests the input and output, including buttons, door locks, lights, and relays.

Variance/Mean – Calculates the variance and mean for each detector.

Scaler – Performs a scaler count for a user-definable count time. Also used to clear fail or contaminated conditions.

FOM – Finds the optimum HV set point for each detector by running a high-voltage plateau and calculating the Figure of Merit.


Efficiencies – Calculates the efficiencies for each detector.

Reset Host – Resets the internal host board. The host board is responsible for collecting the count data and passing it to the SBC through a USB serial port.

View Log – Views the log and report files. System, background, and scan logs are created automatically. Other screens provide the ability to save a report, which can be viewed here.

Section
7

Radiological Setup



The Radiological Setup screen (see further explanation on following pages) provides access to the various parameters and options that control the counting modes, occupancy modes, isotopic mix, background alarm set points, and units of measure. The title indicates the current counting mode.

After making a change, press the Apply button to save the changes immediately. Press the OK button to save the changes and exit the Radiological Setup screen. Press the Close button to exit the screen without saving any changes.

Occupancy Modes

There are four occupancy modes available.

Walk-Thru Mode

Walk-Thru mode allows maximization of personnel throughput, requiring only a short walk-thru scan. As personnel approach the portal, they are sensed by one of the motion sensors and instructed to pass through the portal. Sampling begins once the subject has entered the portal (broken the occupancy beam) and ends shortly after when the subject has passed through the portal. Results of the count are then displayed on the LCDs, as well as indicated by the light stack and

audible voice and noise alarms. Detection sensitivity is limited in this mode by short count times, but personnel can be scanned at a high rate. Counting modes 1-3 and the Fast Alarm/Clean options are not available in the Walk-Thru mode. For more details regarding Walk-Thru mode operation, see Section 5 – Operation.

Pause Mode

Pause mode maintains a simple subject interaction and relatively quick throughput while also providing an extended count time for greater portal sensitivity. As personnel approach the portal, they are instructed to enter and pause inside the portal. Once the portal is entered, the subject is informed that counting has begun. Once the count is complete, results are indicated visually and audibly. For more details regarding Pause mode operation, see Section 5 – Operation.

Pause-And-Turn Mode

Pause-And-Turn mode combines multiple counts with multiple subject positions to improve both scan sensitivity and subject coverage. Personnel are instructed to enter the portal and pause as their approach is detected. Once in the portal, an initial count is taken with the subject standing as if in a doorway. Once this count is complete, the subject is instructed to rotate 90 degrees and again pause as another count is taken. The addition of the second 90-degree rotated count provides for greater exposure of the subject's front and back to the detectors. Upon count completion, visual and audible results are given. For more details regarding Pause-And-Turn mode operation see Section 5 – Operation.

Front-And-Back Mode

Front-And-Back mode minimizes subject distance from detectors by scanning personnel when the front and back of their body is placed in direct contact with the detector covers. As personnel approach, they are instructed to enter the portal. Once inside the portal the subject is instructed to turn 90 degrees (left) and move forward until against the side detectors. A count is then taken followed by instructions to move backward until against the opposite side detectors. Another count is taken and results given visually and audibly. For more details regarding Front-And-Back mode operation, see Section 5 – Operation.

Counting Modes

There are three counting modes available.

Mode 1 – Maximum Sensitivity

The alarm set point is computed at the minimum allowable value that is determined by the background count rate and acceptable false alarm probability.

Note: Alarm set points are computed as background-subtracted (net) count rates.

$$R_{A(min)} = K_B \sqrt{\frac{R_B}{T} + \frac{R_B}{T_B}}$$

Where:

$R_{A(min)}$ = Alarm set point in cps.

K_B = Background sigma coefficient that determines the false alarm probability.

R_B = Average background count rate in cps for the detector.

T = Count time in seconds.

T_B = Background count time in seconds.

After calculating the alarm set point, a check is made to see if the computed MDA exceeds the maximum MDA. If it does, then a high-background alarm is posted.

$$MDA = \frac{R_{A(min)} + K_{S+B} \sqrt{\frac{MDA \times Eff + R_B}{T} + \frac{R_B}{T_B}}}{Eff}$$

Where:

$R_{A(min)}$ = Alarm set point in cps.

K_{S+B} = Composite sigma coefficient for controlling the false alarm probability.

MDA = Minimum Detectable Activity.

Eff = Detector efficiency.

R_B = Average background count rate in cps for the detector.

T = Count time in seconds.

T_B = Background count time in seconds.

Mode 2 – Fixed MDA

The alarm set point is calculated on the basis of the maximum allowable, as determined by the MDA and its associated detection probability.

$$R_{A(\max)} = MDA \times Eff - K_{S+B} \sqrt{\frac{MDA \times Eff + R_B}{T} + \frac{R_B}{T_B}}$$

Where:

$R_{A(\max)}$ = Alarm set point in cps.

K_{S+B} = Composite sigma coefficient for controlling the false alarm probability.

MDA = Minimum Detectable Activity.

Eff = Detector efficiency.

R_B = Average background count rate in cps for the detector.

T = Count time in seconds.

T_B = Background count time in seconds.

If the value of $R_{A(\max)}$ is less than $R_{A(\min)}$, as is computed using the Mode 1 equation, a high-background alarm is posted.

Mode 3 – Minimum Count Time

The minimum count time is calculated based on the user-defined MDA and is rounded up. If the resulting count time is greater than the user-defined maximum count time, then a high background alarm is posted.

$$T_{MIN} = \left[\frac{K_B \sqrt{2R_B} + K_{S+B} \sqrt{MDA \times Eff \times 2R_B}}{MDA \times Eff} \right]^2$$

If the computed value for the minimum count time exceeds the background update time, then the background update time will be extended to match the minimum count time. When this occurs, the status will change to Extending Background.

The minimum count time is calculated as follows:

1. Calculate T_{MIN} using the equation above. If the calculated value of T_{MIN} exceeds the background update time, extend the background to match.
2. If T_{MIN} is less than or equal to the background update time, then the equation below is used to iteratively compute the minimum count time.

$$T_{MIN} = \frac{R_B}{\left[\frac{MDA \times Eff - K_{S+B} \sqrt{\frac{MDA \times Eff + R_B}{T_{MIN}} + \frac{R_B}{T_B}}}{K_B} \right]^2 - \frac{R_B}{T_B}}$$

Where:

T_{min} = Computed minimum count time.

K_{S+B} = Composite sigma coefficient for controlling the false alarm probability.

K_B = Background sigma coefficient for controlling the false alarm probability.

MDA = Minimum Detectable Activity.

Eff = Detector efficiency.

R_B = Average background count rate in cps for the detector.

T_B = Background count time in seconds.

3. The iteration is complete when the resulting T_{MIN} is greater than or equal to the T_{MIN} used on the right side of the equation.
4. If the calculated count time is greater than the user-defined maximum count time, a high-background alarm is posted.
5. The Mode 3 alarm set point is calculated using the $R_{A(MIN)}$ formula for Count Mode 1.

Count Mode Settings

The Settings tab allows the various parameters that affect the operation of the count modes to be modified. After changing a setting, press the Apply button to save the changes immediately.

Note: Only the settings that apply to the current count mode are shown.

The screenshot shows a dialog box titled "Radiological Setup - Mode 1" with four tabs: "Mode", "Settings", "Background Alarms", and "Calculations". The "Settings" tab is active. It contains several settings, each with a text input field and up/down arrow buttons:

- Max MDA (dpm): 55,500
- High Alarm Activity Level (dpm): 100,000
- Count Time (secs): 2
- Detection Probability %: 95.00
- Composite Sigma Coefficient (K sub S+B): 1.65
- False Alarm Probability %: 0.100
- Background Sigma Coefficient (K sub B): 4.00
- Alarm Time Extension Factor: 0

At the bottom of the dialog box are three buttons: "OK", "Cancel", and "Apply".

Max MDA – Sets the Mode 1 maximum MDA value. If the calculated MDA is greater than this value, a high background is posted. This value is set in specified activity units and is automatically converted when changed. The default value is 675 Bq.

High Alarm Activity Level – If an alarm exceeds this value, it is considered to be a high alarm. A high alarm displays a different message than a low-level alarm.

MDA – Sets the minimum detectable activity used in Mode 2 and 3. This value is set in specified activity units and is automatically converted when changed. The default value is 675 Bq.

Count Time – Specifies the count time in seconds for Mode 1 and 2. The count time is adjustable from 6 to 1000 seconds. The default value is 10 seconds.

Max Count Time – Specifies the maximum allowed count time in seconds for Mode 3. If the computed count time is greater than this value, a high

background is posted. The count time is adjustable from 6 to 1000 seconds. The default value is 60 seconds.

Minimum Count Time – Specifies the minimum count time allowed in seconds for Mode 3. The computed count time will not be allowed to be less than this value. The count time is adjustable from 6 to 1000 seconds. The default value is 6 seconds.

Detection Probability % - Sets the detection probability percentage for all counting modes. Changing this value updates the K_{S+B} parameter. The detection probability is adjustable from 0 to 100 percent. The default value is 95%.

K_{S+B} – Sets the composite sigma coefficient for all counting modes. Changing this value updates the Detection Probability % parameter. The default value is 1.65.

False Alarm Probability % - Sets the false alarm probability for all counting modes. Changing this value updates the K_B parameter. The false alarm probability is adjustable from 0 to 100 percent. The default value is 0.10%.

K_B – Sets the background sigma coefficient for all counting modes. Changing this value updates the False Alarm Probability % parameter. The default value is 4.00.

Alarm Time Extension Factor – Sets a multiplier of elapsed time that is used whenever an alarm condition is detected. The alarm time extension factor is adjustable from 0 to 1000. The default value is 0.

Fast Clean

The Fast Clean option provides the ability to determine if personnel are clean before the count time has completed. If enabled, the accumulated count rate is evaluated at the end of each one-second interval throughout the monitoring cycle. If the count rate is statistically shown to have a very high probability of being free of contamination, an early clean condition will be posted immediately after the minimum count time has elapsed. The Fast Clean option only applies to the individual detectors. The following equation is used:

Individual Detectors Enabled

$$\frac{\text{Accumulated Counts}}{T_E} < R_B + MDA \times Eff - K_{S+B} \sqrt{\frac{MDA \times Eff + R_B}{T_E} + \frac{R_B}{T_B}}$$

Where:

T_{\min} = Computed minimum count time.

K_{S+B} = Composite sigma coefficient for controlling the false alarm probability.

MDA = Minimum Detectable Activity.

Eff = Detector efficiency.

R_B = Average background count rate in cps for the detector.

T_B = Background count time in seconds.

T_E = Elapsed count time since the start of the monitoring cycle.

If all single-detector channels are found to satisfy this test, then a clean condition is posted immediately.

Fast Alarm

The Fast Alarm option provides the ability to determine if personnel are contaminated before the count time has completed. If enabled, the counts required for an alarm condition are calculated at the end of each one-second interval and compared to the accumulated counts. A Fast Alarm will only be posted after the minimum count time has elapsed. The following equation is used:

$$\frac{\text{Accumulated Counts}}{T_E} > R_B + K_B \sqrt{\frac{R_B}{T_E} + \frac{R_B}{T_B}}$$

Where:

R_B = Average background count rate in cps for the detector.

K_B = Background sigma coefficient.

T_B = Background count time in seconds.

T_E = Elapsed count time since the start of the monitoring cycle.

Sum zone alarm conditions are not processed using the Fast Alarm algorithm.

Individual Alarms

Alarms can be enabled for individual detectors and sum zones. When the individual detectors are enabled, then the sum zones can also be active.

Units of Measure

The activity units of measure are used where a value is displayed as an activity. When changing the activity unit, the activity is automatically converted into the new unit of measure. The following units of measure are supported:

 dpm

 pCi

 nCi

 μ ci

 mCi

 Ci

 Bq

 kBq

 MBq

The count rate units are used where a value is displayed as a count rate. The following units of measure are supported:

 cps

 cpm

Updating Background

A new background is taken after the Supervisor application is started. The background count time is set here and is valid from 30 to 600 seconds. The default value is 120 seconds. A background update is forced when returning to the Operate screen from the Setup menu, or when the Update button is pressed. During this time the Gamma Personnel Portal is considered not ready for use. Once a new background update is complete, the background is updated once every second.

The Sigma Limit and Sigma OK parameters are used to detect sudden increases or decreases in the background. If the sigma for the current one-second count is greater than the last one-second count, a tentative background update is started. If after 10 seconds the background falls to within the Sigma OK parameters, the background is restored and normal operation continues. If after the 10 seconds background is greater than the Sigma OK parameter, a background update is forced.

Background Update Method

The background update method involves using a simple “sliding window” for computing an average. A FIFO buffer the size of the number of seconds in the background update count time is created. After the full background update, this buffer will be filled with readings in cps. As additional counts are collected, the oldest count will be replaced with the newest count and an average will be calculated using the equation:

$$Avg_{New} = Avg_{Old} + \frac{x_{Newest} - x_{Oldest}}{N}$$

Where:

Avg_{New} = New background average in cps. This is the new value for R_b used in other equations.

Avg_{Old} = Old background average in cps.

x_{Newest} = Newest 1 second count in FIFO buffer.

x_{Oldest} = Oldest 1 second count in FIFO buffer.

N = Number of samples in the FIFO buffer.


To determine when a new background update is required, the following equation is used:

$$\sigma = \frac{x_{Newest} - (Avg_{Old})}{\sqrt{Avg_{Old}}}$$

If the sigma is greater than or equal to Sigma Limit parameter, a tentative 10-second background update is started. If the count rate returns to within the sigma value set by the Sigma OK parameter of the old average rate at the end, the old background average is kept. If it has not, then a background update is forced. The background sigma limit feature can be disabled by unchecking the Enable Sigma Limit checkbox.

Background Alarms

Each detector has a low and high background alarm set point, which is used to identify a detector that is failing. If the background drops below the low set point or goes above the high set point, a detector failure will occur. A low or high background alarm normally signals a failure of either the detector or its associated preamp electronics. A high background alarm can also warn of some nearby strong source of radiation.



Calculations

The Calculations tab shows the current alarm set point and other values relating to the counting mode for all detectors and sum zones. These values are updated once per second. In the far right column, Mode 1 and 3 show the sensitivity, and Mode 2 shows the false alarm probability.

The background is displayed in cpm or cps, depending on the units selected along with the background and alarm set point adjusted for the current count time. The composite efficiency for each detector and sum zone is displayed.

Radiological Setup - Mode 1

Mode Settings Background Alarms Calculations

	Background (cps)	Background (4 samples)	Efficiency	Set Point (4 samples)	Sensitivity (dpm)
1	1,000	400	34.47 %	460	26,111
2	1,000	400	34.47 %	460	26,111
3	1,000	400	34.47 %	460	26,111
4	1,000	400	34.47 %	460	26,111
5	1,000	400	34.47 %	460	26,111
6	1,000	400	34.47 %	460	26,111
7	1,000	400	34.47 %	460	26,111
8	1,000	400	34.47 %	460	26,111
1+2	2,000	800	68.94 %	885	18,463
2+3	2,000	800	68.94 %	885	18,463
3+4	2,000	800	68.94 %	885	18,463
4+5	2,000	800	68.94 %	885	18,463
5+6	2,000	800	68.94 %	885	18,463
6+7	2,000	800	68.94 %	885	18,463
7+8	2,000	800	68.94 %	885	18,463
8+1	2,000	800	68.94 %	885	18,463

^

v


OK Cancel Apply

Section

8

Operational Setup

The Operational Setup screen is used to configure the various options that define how the Supervisor software operates.



General

Serial Number

This sets the serial number of the Gamma Personnel Portal. This is displayed on the title bar of the Operate screen. The serial number is also stamped onto the frame of the Gamma Personnel Portal near the AC Power input panel.

Site

This user-defined field is 20 characters in length. The Site is displayed in the title bar of the Operate screen.

Location

This user-defined field is 15 characters in length. The Location is displayed in the title bar of the Operate screen.

Customer ID

This user-defined field is 10 characters in length. The Customer ID is displayed in the title bar of the Operate screen.

Password

Setup involves a two-level numeric password. A level 1 password does not allow access to the following setup screens: Radiological, Operational, Detectors, FOM, Efficiencies, Variance/Mean, and False Alarm. A password is required to access the setup menu and to exit/reboot/shut down the application.

Alarm Hold

This function sets the number of seconds before the alarm will automatically clear. When the alarm hold time is set to zero, pressing the Acknowledge button once within six seconds of the alarm will silence the audio, and pressing it again after six seconds will clear the alarm. When the alarm hold time is greater than zero, pressing once will silence the audio, and pressing again will have no effect.

Infrared Delay

The infrared delay controls how long background resumes updating after the portal is cleared.

Infrared Timeout

The infrared timeout triggers a clear portal warning if the person does not walk through the portal. This can indicate a problem with one of the infrared sensors.

Enable Check-out Mode

The Enable Check-out mode option allows infrared sensor inputs to be ignored for the purpose of testing, calibration and maintenance.

Options

The screenshot shows a dialog box titled "Operational Setup" with a tabbed interface. The "Options" tab is selected. The dialog contains nine settings, each with a radio button for "Yes" and "No". The settings are:

Setting	Yes	No
Require employee ID to start count	<input type="radio"/>	<input checked="" type="radio"/>
Require Password to Clear Alarms	<input type="radio"/>	<input checked="" type="radio"/>
Enable Automatic Recount	<input type="radio"/>	<input checked="" type="radio"/>
Enable Incomplete Alarm	<input type="radio"/>	<input checked="" type="radio"/>
Show Alarm Result	<input checked="" type="radio"/>	<input type="radio"/>
Show Counts as Activity	<input checked="" type="radio"/>	<input type="radio"/>
Voltage Cal Enabled	<input type="radio"/>	<input checked="" type="radio"/>
Latch Failures	<input type="radio"/>	<input checked="" type="radio"/>
Ignore Approach/Exit	<input checked="" type="radio"/>	<input type="radio"/>

At the bottom of the dialog are three buttons: "OK", "Cancel", and "Apply".

Require Employee ID to Start Count

When enabled, the user must enter in an Employee ID before a monitoring cycle can be started. The employee ID is logged in the scan log.

Require Password to Clear Alarms

When enabled, the level 1 or 2 password is required before the alarm can be cleared.

Enable Automatic Recount

Automatic recount causes a second count to be taken immediately after an alarm count.

Enable Incomplete Alarm

When enabled, an incomplete alarm is triggered anytime the subject leaves the portal before completing the count procedure.

Show Alarm Result

When enabled, this option will display as a percentage of how much over the alarm set point the reading was.

Show Counts as Activity

When enabled, this option will display the counts during occupancy in activity units instead of count rate units.

Voltage Cal Enabled

When enabled, the voltage calibration button will be shown on the detectors screen.


Latch Failures

When enabled, this option will require user intervention to clear any failures. When disabled, the instrument will return to READY status automatically if the failure condition goes away. This is useful when high background failures will occur due to known activity throughout the day.

Ignore Approach/Exit Sensors


When set to YES, the portal will ignore the approach and exit infrared motion sensors for triggering a counting sequence. The sensors will still be used in clearing the portal after an occupancy and for maintaining distance around the portal for a new background update.

Language




The Language tab allows the various audible voice alerts to be individually selected. The voice alert files are standard WAV files and can be replaced with custom audio files.

Logging




The Logging screen is used to set whether data logging is enabled, the directory where the data will be logged, and how often to log the background readings. When logging is disabled, no data will be saved to the scan, background, and system event logs.



Com Port

The COM Port screen is used to select the RS-232 communication port used by the onboard computer to the host board. This is set at the factory and normally should not need to be changed.



Network

The Network screen is used to configure the network IP address to a server running the LMI Universal software (Part #: 1370-093).

Section

9

Detectors

The Detectors screen is used to view and set the detector's high voltage and lower level discriminator. The calibration constants can be set through an easy-to-use wizard, which guides the user through the process. The current background readings can also be viewed here.

Background

The background screen (see screen shot on following page) shows the current background detail for each detector. Status for each detector is displayed using the background color of the detector indicators. Count data is received from the host board every 100 ms. These 100 ms counts are accumulated together every one second. The Average count is the current background average. The Sigma is the standard deviation from the old background average and the new one-second count.

These values only update when the background is updating. The detector indicators will change color to indicate the detector's status with green for OK, yellow for fail, and red for alarm. The current Gamma Personnel Portal status is displayed at the bottom. Press the Update button to start a new background update cycle.

Detectors

Background	HV	LLD	S/N	
	100 ms	1 Sec	Average (cps)	Sigma
1	100	1000	1,000	0.00
2	100	1000	1,000	0.00
3	100	1000	1,000	0.00
4	100	1000	1,000	0.00
5	100	1000	1,000	0.00
6	100	1000	1,000	0.00
7	100	1000	1,000	0.00
8	100	1000	1,000	0.00


Current Status:

High Voltage (HV)

	Set Point (V)	Readback	Actual CC	Readback CC
1	1,000	0	+0.0	+0.0
2	1,000	0	+0.0	+0.0
3	1,000	0	+0.0	+0.0
4	1,000	0	+0.0	+0.0
5	1,000	0	+0.0	+0.0
6	1,000	0	+0.0	+0.0
7	1,000	0	+0.0	+0.0
8	1000	0	+0.0	+0.0

The HV screen displays the current high-voltage set point and calibration constants for each detector. High voltage is typically set using the FOM-HV screen. Read-back voltage is the value from the A/D converter. There are two calibration constants associated with the high voltage. The Actual calibration constant is used to calibrate the high-voltage output while the read-back calibration constant is used to calibrate the read-back value from the A/D converter. The high voltage is adjustable from 0 to 2500 V and the calibration constants are adjustable from -9.9 to +9.9. Click the Refresh button to reload the parameters from all detector boards. After changing the high-voltage set point, click the Save button to apply the changes. The calibration constants can only be changed from the Voltage Calibration Wizard. See Page 41 for more information on the Voltage Cal button.

Lower Level Discriminator (LLD)



The LLD screen displays the current lower-level set point and calibration constant for each detector. Read-back is the value from the A/D converter, and should be close to the set point. The Actual calibration constant is used to calibrate the LLD output. The LLD is adjustable from 0 to 3300 mV, and the calibration constant is adjustable from -9.9 to +9.9. Click the Refresh button to reload the parameters from all detector boards. After changing the LLD set point, click the Save button to apply the changes. The calibration constant can only be changed from the Voltage Calibration Wizard. See the following page for more information on the Voltage button.


Voltage Calibration

The Voltage Calibration Wizard (see next page) is accessed through the Voltage button. The wizard guides the user through the process of setting the calibration constants for each detector. The following parameters will be set by the wizard:

- ☢ HV Actual Calibration Constant

- ☢ HV Read Back Calibration Constant
- ☢ LLD Actual Calibration Constant

The wizard automatically calculates the calibration constant based on the measurements entered for each detector. All calibration constants are zeroed out at the beginning. Calibration requires a Model 500 Pulser with high-voltage readout or a high-impedance voltmeter, with at least 1000 Megohm meter input resistance. Click the Next button to begin the calibration.



HV Actual Calibration Constant

Click the Begin button to start the process of calibrating the HV Actual Calibration Constant. The calibration constant is calculated automatically by entering the high voltage measured at the detector connector. The wizard will set the high voltage of all detectors to zero and then prompt for the cable to be connected to the first detector. The detector order is: 1, 2, 3..., 7, 8.

Voltage Calibration Wizard

Step 1: HV Actual Cal Constant

	Set Point	Measured	Cal Constant
1	0	0	+0.0
2	0	0	+0.0
3	0	0	+0.0
4	0	0	+0.0
5	0	0	+0.0
6	0	0	+0.0
7	0	0	+0.0

This step will calibrate the HV measured from the detector connector.

Measure the HV from the detector connector of each channel and enter the values in the fields to the left.


The software will calculate the correct calibration constant automatically. Click Begin to measure the HV at each connector.

When finished click Next to continue or Cancel to exit.

HV Actual Calibration


Connect the cable to Detector 1 and then click Next to continue.

After connecting the cable and clicking the Next button, the high voltage will be restored to the set point. Enter the measured high voltage and click Next. Continue repeating this process until all detectors are completed.



HV Read Back Calibration Constant

The calibration of the Read Back value is automatic. The wizard will automatically calculate a new calibration constant each time the Reload button is clicked. If the values returned are consistent, click the Next button.



LLD Actual Calibration Constant

The LLD actual calibration constant is obtained by measuring the LLD from the test point on each detector and entering the measured reading in the appropriate field. The calibration constant is automatically calculated. Click the Next button when finished.

Voltage Calibration Wizard

Step 3: LLD Actual Cal Constant

	Set Point	Measured	Cal Constant
1	0	0	+0.0
2	0	0	+0.0
3	0	0	+0.0
4	0	0	+0.0
5	0	0	+0.0
6	0	0	+0.0
7	0	0	+0.0

This step will calibrate the LLD measured from the LLD Test-point.

Measure the LLD from the test-point of each channel and enter the values in the fields to the left.

The software will calculate the correct calibration constant automatically.

When finished click Next to continue or Cancel to exit.

Detector Serial Numbers

The detector serial numbers screen allows the user to enter and save the serial number of each detector in the Model 53 Portal Monitor. This is to allow easy look-up and tracking of detector numbers via software, rather than being forced to open the Model 53 covers in order to check detector serial numbers.

Detectors

Background	HV	LLD	S/N
1	<input type="text"/>	<input type="text"/>	<input type="text"/>
2	<input type="text"/>	<input type="text"/>	<input type="text"/>
3	<input type="text"/>	<input type="text"/>	<input type="text"/>
4	<input type="text"/>	<input type="text"/>	<input type="text"/>
5	<input type="text"/>	<input type="text"/>	<input type="text"/>
6	<input type="text"/>	<input type="text"/>	<input type="text"/>
7	<input type="text"/>	<input type="text"/>	<input type="text"/>
8	<input type="text"/>	<input type="text"/>	<input type="text"/>

Section

10

Source Check

The Source Check screen (see screen shot on following page) can be used to perform a daily test on the detectors to verify correct function.

Enter the source size of the check source and a percentage that the activity must fall within to be considered acceptable. This percentage is applied to the source activity as well as the background estimate since background fluctuations can cause a significant activity estimate swing, particularly when the source is relatively small in relation to the background. Select an Isotope from the isotope library in the Isotope Efficiency drop-down list. The collected efficiency data of this isotope is used to estimate the source activity during a source check. Select the location from which the efficiency data for the selected isotope was collected, either from the portal center line, 3 inch, or detector cover surface positions. This information is used to apply a correction factor to the isotope efficiency data that allows source checks to be performed on the detector cover surface at marked locations on the detector covers even if the calibration was performed along the detector centerline (which is the recommended calibration procedure).

Click the Start button to start the test. Position the source on the detector surface directly over the cover marking (a small dimple in the metal) of each detector in any sequence desired. The system will run continuous counts using the normal monitoring count time until all detectors pass or the test is cancelled.

Source Check

Source Activity:

Pass Percent ±:

Isotope Efficiency:

Efficiency Location:
 Center Line 3 Inch Surface

Detector	Activity (dpm)	Status	Background (dpm)
1	<input type="text" value="0"/>		<input type="text" value="0"/>
2	<input type="text" value="0"/>		<input type="text" value="0"/>
3	<input type="text" value="0"/>		<input type="text" value="0"/>
4	<input type="text" value="0"/>		<input type="text" value="0"/>
5	<input type="text" value="0"/>		<input type="text" value="0"/>
6	<input type="text" value="0"/>		<input type="text" value="0"/>
7	<input type="text" value="0"/>		<input type="text" value="0"/>
8	<input type="text" value="0"/>		<input type="text" value="0"/>

Instructions:

1. Position source for first detector test.
2. Click the start button. This will start a continuous count cycle.
3. Move source to each detector that requires testing.

Once all detectors have passed the system is operational.
If any detector can not pass the test, service is required.

Section

11

False Alarm

The False Alarm screen (see screen shot on following page) is used to run a series of counts to determine the false alarm rate. The number of samples to run defaults to 1000, but can be changed from 10 to 100,000. The Gamma Personnel Portals current background update and count times are used to perform the test.

After setting the number of samples, click the Start button to begin the false alarm test. A new background will be acquired before the test is started. The current sample number, count time remaining, end time, and total false alarms are displayed.

Should an alarm be posted, an immediate background update will follow so as to track any potentially changing background average.

During the false alarm test, the current background, gross count, total alarms, alarm set point, and highest count are displayed in the grid for each detector. Click the Cancel button to stop the test.

The criterion for an alarm is satisfied when the net (background-subtracted) count rate for any detector exceeds the alarm set point (R_A).

$$R_A = K_B \sqrt{\frac{R_B}{T} + \frac{R_B}{T_B}}$$

Where:

R_A = Alarm set point in cps.

K_B = Background sigma coefficient, which determines the false alarm probability.

R_B = Average background count rate in cps for the detector.

T = Count time in seconds.

T_B = Background count time in seconds.

Only one false alarm will be reported regardless of how many detectors alarmed.

Test False Alarm

Settings

Number of Samples ▲ ▼

Test

Sample Number Count Time Remaining

Start Time End Time

Total False Alarms

Multiple detector alarms for a sample are counted as 1 false alarm

Detector	Background (cps)	Count	Total Alarms	Alarm Set	Highest Count
1		0	0		0
2		0	0		0
3		0	0		0
4		0	0		0
5		0	0		0
6		0	0		0
7		0	0		0
8		0	0		0
1+2		0	0		0
2+3		0	0		0
3+4		0	0		0
4+5		0	0		0
5+6		0	0		0
6+7		0	0		0
7+8		0	0		0
8+1		0	0		0
8+2+1		0	0		0


Section

12

Buttons/Relays

The Buttons/Relays screen (see screen shot on following page) is used to test the various inputs and outputs on the Gamma Personnel Portal.

The outputs include the lights in the stack and relays. The inputs include the IR sensors, occupancy sensor, and pushbuttons. Click on the buttons to turn on/off the lights/relays. When a pushbutton is pressed, the indicator lights up to indicate the button was read successfully. The Gamma Personnel Portal should be taken out of service if any button, sensor, or light fails to respond correctly.



Section

13

Variance/Mean

The Variance/Mean screen (see screen shot on the following page) is used to determine the stability of the background and should be run after the Gamma Personnel Portal is set up. The number of samples and count time are user adjustable.

Click the Start button to begin the test. The current sample number, count time remaining, and end time are displayed along with the detector count, mean, and variance. Click the Cancel button to stop the test.

The algorithm for computing the variance and mean at the end of each sampling period follows:

Initialization at beginning of Variance/Mean Test

```
mean = 0
sampleCounter = 0
m2 = 0
```

At end of each count:

```
sampleCounter = sampleCounter + 1
delta = count - mean
mean = mean + delta
m2 = m2 + delta * (reading / sampleCounter)
variance = m2 / (sampleCounter -1)
```

Test Variance/Mean

Settings Start

Number of Samples ▲ ▼

Count Time (seconds) ▲ ▼

Test

Sample Number Count Time Remaining Start Time

End Time

Detector	Current Count	Variance	Mean	Ratio V/M
1	0	0	0	0
2	0	0	0	0
3	0	0	0	0
4	0	0	0	0
5	0	0	0	0
6	0	0	0	0
7	0	0	0	0
8	0	0	0	0


Save Report Close

Section

14


Scaler

The Scaler screen provides a way to start a count and see the results as gross counts and count rate. It is also used to clear System Failure conditions.



Set the desired count time in seconds and click the Start button. Click the Cancel button to stop a count that is in progress.

If the Gamma Personnel Portal is in a fail condition caused either by a background alarm or contamination, a fourth column will be displayed showing the last good background for comparison along with a button that will be used to clear the fail condition.



Once the contamination has been cleaned off the detectors or the cause of the fail has been corrected, take a scaler count to verify that the background values have returned to normal. If the background has returned to normal for all detectors, click Clear Fail to clear the fail condition and exit back to the main screen. A residual contamination check will be automatically performed followed by a background update if no contamination is detected.


Section

15

FOM-HV

The FOM-HV screen is used to calculate a Figure of Merit to find the optimum high voltage operating point for each detector. A background count and source count are taken at various high-voltage set points, and the best high voltage is chosen.

The Settings tab (see following page) allows choosing of the detectors that are calibrated, along with the count time for background and source counts.



Select Detectors for HV Plateau

Select which detectors will be included in the calibration. The FOM-HV calibration can be run on multiple detectors or just a single detector.

Count Time Settings

The background and source count times can be set individually and are adjustable from 1 to 300 seconds.

Source Size


The source size in DPM is used to calculate the detector efficiency.

HV Settings

The start and end of the high-voltage plateau are defined. The default start high voltage is 700 V, and the default end high voltage is 1100 V. The high-voltage increment defaults to 25 V and can be adjusted from 1 to 500 V.

Running the FOM-HV

Click the Start button to begin the FOM process (see screen shot on the following page). The FOM tab will automatically be selected if necessary. This tab shows a grid that displays the background, source, and net counts, as well as the FOM and efficiency at each high-voltage step.



A prompt will be displayed indicating that all sources should be removed from the chamber. After acknowledging this prompt, background counts will be taken at each high-voltage step.

When all background counts are complete, a prompt will be displayed indicating that the source must be placed in position for the detector(s) chosen. This is done for each detector selected. Place the source appropriately (see the Model 53 Calibration procedure for more detail). Click the OK button when done.

The Gamma Personnel Portal will begin taking source counts at each high-voltage step until completed. Once the current detector is complete, repeat the process for all detectors desired.

When the FOM is complete, the recommended voltage will be displayed. Click the row of the desired HV value. Click the Set HV button to set the detector's high voltage to the selected row set point.

The Figure of Merit is calculated using the following equation:

$$FOM = \frac{S^2}{(\sqrt{B} + \sqrt{S + B})^2}$$

Where:

S = Net (background subtracted) source count rate

B = Background count rate

Click the Save button to save an FOM report. The FOM-HV results are also displayed as a graph of the three curves. The curves are normalized, i.e., each one peaking at 1, since the scales of the curves can vary dramatically from one another. The background curve is displayed as blue, the net source curve is displayed as red, and the FOM curve is displayed as green.

Section


16

Efficiencies

The Efficiency screen is used to configure the efficiencies for each detector. These efficiencies are used in calculating alarm points and determining the specific activity if the weight option is enabled.

The Configuration tab (see the following page) is where the count times and detectors are selected. In addition to the individual detectors, the Sum channel can also be selected. The count times are adjustable from 1 to 300 seconds.

Configuration



The Sources tab is used to configure up to five sources that will be used in determining the efficiencies.

Sources

The screenshot shows a software window titled "Calibrate Efficiencies" with four tabs: "Configuration", "Sources", "Isotopic Mix", and "Efficiencies". The "Sources" tab is active. On the left, a list contains "Source-1" (highlighted), "Source-2", "Source-3", "Source-4", and "Source-5". Below the list is a "Save" button. The main area contains the following fields:

- Isotope: Source-1
- Half-life (days): 10,000 (with up/down arrows)
- Certification Date: 01/01/2009 (with a calendar icon)
- Activity: 1,000,000 (with up/down arrows) and a unit dropdown set to "dpm"
- Current Activity: 8.4807E+005 (with a unit dropdown set to "dpm")

At the bottom, there are "Save Report" and "Close" buttons, and a status bar showing "Time Remaining: 00:00".

Up to five sources can be used in determining the efficiency. For each source, the half-life, certification date, and activity can be set. The current activity is automatically determined based on the data entered for the source and the current system time.

To edit a source, select it from the list. The default names for the sources are Source-1, Source-2, Source-3, Source-4, and Source-5. After selecting a source, its parameters can be changed. After changing the parameters, click the Save button to save the changes.

The current activity is calculated using the following equation:

$$A_F = A_O e^{-KT}$$

Where:

A_F = Current activity

A_O = Original activity

K = $0.693 /$ half-life in days

T = Elapsed time in days from certification date

Click on the Efficiencies tab to view efficiencies for the detectors.

Isotopic Mix

The screenshot shows a software window titled "Calibrate Efficiencies" with four tabs: "Configuration", "Sources", "Isotopic Mix", and "Efficiencies". The "Isotopic Mix" tab is active. Inside the window, a message states "The sum percentage of all sources must add up to 100". Below this, there are five rows, each with a source name and a percentage value in a text box with up/down arrows: Source-1 (100), Source-2 (0), Source-3 (0), Source-4 (0), and Source-5 (0). A "Total is 100%" message is displayed below the source entries. At the bottom of the window, there are two buttons: "Save Report" and "Close". A status bar at the bottom right shows "Time Remaining: 00:00".

The isotopic mix is used to create a weighted average efficiency during monitoring cycles. Up to five isotopes may be used, which are defined in the Sources tab. The percentages must add up to 100%. The isotope names default to Source-1, Source-2, Source-3, Source-4, and Source-5 but will be updated to reflect the isotope names defined in the Sources tab.

Efficiencies

Calibrate Efficiencies

Configuration Sources Isotopic Mix **Efficiencies**

Select Isotope
Source-1

Detector	Background Count	Gross Count	Net Count (cpm)	Efficiency
1	600	30000	294000	34.47 %
2	600	30000	294000	34.47 %
3	600	30000	294000	34.47 %
4	600	30000	294000	34.47 %
5	600	30000	294000	34.47 %
6	600	30000	294000	34.47 %
7	600	30000	294000	34.47 %
8	600	30000	294000	34.47 %

Time Remaining: 00:00

A grid displays the count and efficiency data for all detectors. Each detector has efficiency data for five isotopes for a total of 5 separate efficiencies.

The efficiency is updated only for the selected combination of detector(s) and isotope. After selecting the detectors and isotope, click the Start button to begin. A background will be taken for all selected detectors. After the background count is complete, place the source in the appropriate location for each detector. The efficiency will be calculated after the source count is taken and is the net count divided by the source size.

Composite Efficiency

The composite efficiency is the sum of the isotopic efficiencies multiplied by their isotopic mix percentages.

Section

17

View Logs

The View Log screen is used to view various log files and reports. Log files and reports are saved in the Data Directory specified in the Logging tab of the Operational screen. The Gamma Personnel Portal creates several log files automatically. These are:

- ☢ System Log
- ☢ Background Log
- ☢ Scan Log

Log files are named with a prefix of the date and time in the format of YYYYMMDD and have a file extension of .log. These files are ASCII text files and are viewable in any program that can read text files. New log files are created for each day.

The system log file records events relating to the status of the Gamma Personnel Portal such as:

- ☢ Starting and stopping the Supervisor application
- ☢ Commands sent to the host board
- ☢ Current status
- ☢ Push button events
- ☢ Alarms and failures

The background log file records the current background and detector status at an interval specified in the logging tab of the Operational screen.

The scan log file records information about the results of the monitoring of personnel, including:

- ☢ Date/Time

- ☢ Employee ID, if required
- ☢ Status and which detector alarmed, if any
- ☢ Count time
- ☢ Detector readings

Most setup screens have an option to save a report. These reports can be viewed here as well.

Section

18

Changes

Firmware Compatibility

Version 1.4.0 and below - 21505N01

Version 1.4.2

Fixed issue where non-latching alarms didn't recover when the background increased above the sigma limit.

Version 1.4.1

Updated so forms automatically resize to the current monitor resolution.

Version 1.4.0

Default to COM3 initially.

Fixed where COUNTING was displayed using yellow background in Walk-Thru mode.

Fixed issue displaying 1 second count in detectors screen with a detector failure.

Fixed issues with handling of infrared sensors, which could lead to a failure to clear portal correctly.

Fixed so infrared delay parameters cannot be set below 1.

Fixed issue where background count time could continue past zero.

Version 1.3.9

Minor bug fix.

Version 1.3.8

Added non-latching fail option.

Version 1.3.7

Fixed issue with background update time.

Version 1.3.6

Allow background update time to go as low as five seconds.

Fixed issues with scan log.

Version 1.3.5

Changed background update algorithm for better averaging.
Added option to ignore approach sensors.

Version 1.3.4

Fixed issues with scan log data not showing highest value.
Added more decimal places on source check screen.
Fixed issues with voltage cal.

Version 1.3.3

Fixed issue where sum zone alarm 6+7+8 was reported as 4+5+6 in summary box. Detector indicators were correct.

Version 1.3.2

Added option to disable background sigma limit.

Version 1.3.1

Fixed where Source Check filter shows all reports in the log
Fix for display problem when activity units set to curies

Version 1.3.0

Minor bug fixes and updated default values.

Version 1.2.2

Added fixes from version 1.1.3

Version 1.2.1

Fixed issues with LMI Universal software

Version 1.2.0

Added support for LMI Universal software

Version 1.1.3

Fixed issue where sum alarms were still being checked even though they were disabled.

Removed FTDI driver from installer

Removed sum channel from reports and screens

Version 1.1.2

Fixed missing detector 3 and 6 FOM data in the calibration report

Version 1.1.1

Minor bug fixes

Version 1.1.0

First purchased unit release.

Version 1.0.0

First release.

Section

19

Recycling

Ludlum Measurements, Inc. supports the recycling of the electronic products it produces for the purpose of protecting the environment and to comply with all regional, national, and international agencies that promote economically and environmentally sustainable recycling systems. To this end, Ludlum Measurements, Inc. strives to supply the consumer of its goods with information regarding reuse and recycling of the many different types of materials used in its products. With many different agencies – public and private – involved in this pursuit, it becomes evident that a myriad of methods can be used in the process of recycling. Therefore, Ludlum Measurements, Inc. does not suggest one particular method over another, but simply desires to inform its consumers of the range of recyclable materials present in its products, so that the user will have flexibility in following all local and federal laws.

The following types of recyclable materials are present in Ludlum Measurements, Inc. electronic products, and should be recycled separately. The list is not all-inclusive, nor does it suggest that all materials are present in each piece of equipment:

- ☢ Batteries
- ☢ Glass
- ☢ Aluminum and Stainless Steel
- ☢ Circuit Boards
- ☢ Plastics
- ☢ Liquid Crystal Display (LCD)



Ludlum Measurements, Inc. products, which have been placed on the market after August 13, 2005, have been labeled with a symbol recognized internationally as the “crossed-out wheelie bin.” This notifies the consumer that the product is not to be mixed with unsorted municipal waste when discarding; each material must be separated. The symbol will be placed near the AC

receptacle, except for portable equipment where it will be placed on the battery lid.



REVISION HISTORY



REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP Jew	DATE 1-13-23
DWG NUM: 4540-1159			SCALE: 1/15	
TITLE: M 53 ASSEMBLY-STEP 1 (2022)				
LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159	

REVISION HISTORY


REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC




DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP JOW	DATE 1-12-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53 ASSEMBLY-STEP 2 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159A	

REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



M 53 PROXIMITY IR ASSEMBLY
4540-939
SEE SHEET 540 X 939 (A)






DETAIL B
SCALE 1/2

DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP JW	DATE 1-12-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53-ASSEMBLY-STEP 3 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159B	

REVISION HISTORY


REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC




DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP J6W	DATE 1-13-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53 ASSEMBLY-STEP 4 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159C	

REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC





DETAIL A
SCALE 1 / 7



DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP JGW	DATE 1-12-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53 ASSEMBLY-STEP 5 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159D	

REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



4540-1178
(EMITTER)

20-9187
4540-1178
(RECEIVER)


DETAIL A
SCALE 1 / 3

20-9187



DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP JGw	DATE 1-12-23
DWG NUM: 4540-1159			SCALE: 1/20	
TITLE: M 53 ASSEMBLY-STEP 6 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159E	

REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



DETAIL B
SCALE 1 / 5




DETAIL A
SCALE 1 / 4

DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP Jow	DATE 1-13-23
DWG NUM: 4540-1159			SCALE: 1/20	
TITLE: M 53 ASSEMBLY-STEP 7 (2022)				
LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159F	



REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC

DETAIL A
SCALE 1/6



DETAIL B
SCALE 1/6





ELECTRONICS

DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP Jaw	DATE 1-12-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53 ASSEMBLY-STEP 8 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159G	


REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC




DETAIL A
SCALE 1 / 4
BOTH SIDES


DWN	DATE	DESIGN AUTHORITY	APP	DATE
CMC	1-12-23	JWITT	Jew	1-13-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53 ASSEMBLY-STEP 9 (2022)				
LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES	SHEET	
		540	1159H	





REVISION HISTORY			
REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP New	DATE 1-13-23
DWG NUM: 4540-1159			SCALE: 1/20	
TITLE: M 53 ASSEMBLY-STEP 10 (2022)				
 LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159I	




REVISION HISTORY			
REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



DWN	DATE	DESIGN AUTHORITY	APP	DATE
CMC	1-12-23	JWITT	lw	1-12-23
DWG NUM: 4540-1159			SCALE: 1/12	
TITLE: M 53 ASSEMBLY-STEP 11 (2022)				
LUDLUM MEASUREMENTS, INC. 501 OAK STREET SWEETWATER, TEXAS 79556		SERIES	SHEET	
		540	1159J	

REVISION HISTORY

REV	DESCRIPTION	DATE	BY
1	VALID	1-12-23	CMC



DWN CMC	DATE 1-12-23	DESIGN AUTHORITY JWITT	APP JW	DATE 1-12-23
DWG NUM:4540-1159			SCALE: 1/12	
TITLE:M 53 ASSEMBLY-STEP 12 (2022)				
 LUDLUM MEASUREMENTS, INC. 301 OAK STREET SWEETWATER, TEXAS 79556		SERIES 540	SHEET 1159K	